

DrägerService

Dräger Safety Pacific

Unit 4, 24 Bishop Dunn Place

East Tamaki

Auckland

Tel : (09) 271 4691

Fax : (09) 271-4695

TEST CERTIFICATE

Owner : **CHCh Pest Control
Bromley ChCh**

Docket No: **1685**

Instrument	Part No:	Serial No:	last serviced	next service
GA Quatro	GAQF-03_000	QA 113-015568	29/05/2017	November-17
Tag # NONE				

Sensor :	Combustible	Oxygen	Toxic 1	Toxic 2
Type :	LEL	O ₂	CO	H ₂ S
Units :	Vol% CH ₄	Vol%	ppm	ppm
Serial No :				
Post cal.zero	0.0	20.9	0.0	0.0
Post cal.span	50.0	18.0	100.0	15.0
Cal.gas conc.	50.0	18.0	100.0	15.0
Alarm 1	5.00	19.5	25.0	10.0
Alarm 2	10.00	23.5	200.0	15.0
HYGIENE MODES				
TWA			25.0	10.0
STEL			50.0	15.0
CalGas Lot/Exp	Lot No WO126312-09		Exp 10/2018	

Inspection Summary:

Visual:	OK
Functional:	OK
Calibration verified:	OK

Comments :

All system working normally date and time reset.

Technician :



G.V.Bates.....